

INSTRUCTION BOOKLET

Thank you for choosing our Trackfresh 60 Evo device. This device has been designed to monitor the goods temperature subject to compliance with the cold chain during the transport. For a correct operation we advise you to read carefully the instructions below:

TECHNICAL FEATURES - TRACKFRESH 60 EVO	
TYPE	PRE-PROGRAMMED/ SINGLE USE
MEMORY CAPACITY	17280 READINGS
BATTERY	3.0V CR2032 LITIUM BATTERY
SHELF LIFE	18 MONTHS
START MODE	MANUAL START
BUTTON	START/STOP
LOGGING INTERNAL	5 MINUTES
ALLARM RANGES	ACCORDING TO COSTUMERS NEEDS
ALLARM TYPE	ACCORDING TO COSTUMERS NEEDS
RUN TIME	UP TO 60 DAYS
DISPLAY	LED (OK, ALLARM; FAULT)
SOFTWARE	NO SOFTWARE REQUIRE
AUTOMATIC DATA FILE	PDF FORMAT
DATA SPECIFICATION	TEMPERATURE /DATA/TIME
CONNECTION	USB 2.0 CONNECTOR
COMPATIBLE OS	WINDOWS XP/7/8/10/ ANDROID
PROTECTION	IP 67/NEMA 6
CERTIFICATION	CE/EN12830/RoSH
TEMPERATURE MONITORING	
SENSOR	INTERNAL NTC
MEASUREMENT	-30° ; +70°
ACCURACY	+/- 0,5°
PACKING INFORMATION	
DIMENSION	76 MM X 31 MM X 4,5 MM
WEIGHT	15 GR.
PACKING MATERIALS	FOOD/SAFETY PLASTIC BAG / NYLON+PE
SPQ	20 PCS



www.trackfresh.eu



01

Press and hold the START / STOP button for 5 seconds to start recording. The green LED will flash 5 times , has started normally. Than remove the serial number stickers and attach them to the transport documents or invoice, and place the device in the cargo or wherever you consider better to record.



02

Press quickly once, if you want to check the operating status of the device. If it is recording, the green LED will flash, which indicates that the recording is proceeding correctly.



03

Press the START / STOP button for 5 seconds to stop the recording, the red LED will flash 3 times. This indicates that the recording is successfully concluded.



04

After finishing the recording, take out the device from the transparent plastic case and insert the interface into the USB port of your PC. The device will automatically generate the report in PDF without installing any software.

IMPORTANT

- 1) In the case the device is ON, and the orange LED flashes, it will indicate an operating anomaly. We recommend that you do not use that device, it will be replaced under warranty in the next order.
- 2) Be very careful when turning on the device, avoid double clicking on the START / STOP button, the device may switch on and off at the same time. If this state occurs, immediately stop pressing the button and wait a few minutes before resuming the ignition phase.
- 3) Leave the device sealed in the case, it will allow you to use it even in conditions of high humidity or in water.

How it is used



1) Push START/STOP button to start



2) Put the started Trackfresh 60 evo to anywhere of your shipping



3) Connect to PC by USB port



4) Directly get the PDF report automatically